

Technical Compendium

# CDI OneView™ Monitoring System

An overview of the CDI OneView Monitoring System and its industry leading technology.



CE-Mark Pending


**TERUMO**  
CARDIOVASCULAR

## The CDI OneView Monitoring System Displays up to 22 Vital Patient Parameters.

The CDI OneView Monitoring System was designed and developed to enable continuous monitoring of in-line patient parameters during cardiopulmonary bypass (CPB) — potential of hydrogen (pH), partial pressure of carbon dioxide (pCO<sub>2</sub>), partial pressure of oxygen (pO<sub>2</sub>), potassium ion (K<sup>+</sup>), temperature, oxygen saturation (SO<sub>2</sub>), hematocrit (HCT), and hemoglobin (Hgb). In addition, the system offers continuous base excess (BE), bicarbonate (HCO<sub>3</sub><sup>-</sup>), oxygen consumption (VO<sub>2</sub>), oxygen delivery (DO<sub>2</sub>), oxygen extraction ratio (O<sub>2</sub>ER), cardiac index (CI), measured flow, visibility of cerebral saturations, and area under the DO<sub>2i</sub> curve (AUC).

Using optical fluorescence and reflectance technologies with disposable Shunt Sensors and Cuvettes placed in the extracorporeal circuit, the CDI OneView Monitoring System monitors and displays real-time changes in patient parameters. The system provides continuous results and reduces the need to obtain or dispose of blood samples as required with laboratory analyzers. Adoption of CDI OneView Monitoring System technology allows constant visibility and provides early warning of dynamic changes.

By providing a comprehensive view of all relevant information, the CDI OneView Monitoring System enables clinicians to gain valuable insights, make informed decisions, and drive strategic outcomes.



**Figure 1:** The CDI OneView Monitoring System Shunt Sensors and H/S Cuvettes are placed in the extracorporeal circuit, allowing real-time response to changes in patient parameters.

## System Accuracy Limits

The CDI OneView Monitoring System has been subjected to rigorous bench tests to simulate clinical use and assess accuracy and precision over the system operating ranges. Blood samples taken from the test circuit were analyzed by conventional analyzer on a sample-by-sample basis and compared to values displayed by the CDI OneView Monitoring System. The absolute values of the differences were then used to calculate error for each measurement.

$$Error = | CDI - Analyzer |$$

The population of errors for each sensor is summarized statistically using a Weibull Distribution.

**Note:** This analysis method using absolute value utilizes a different statistical approach from legacy CDI System products that is more suitable for the distribution of data across all operating conditions. Therefore, the statistical summary should not be used for direct comparison to legacy CDI System product performance specifications.

Sensor	Median Error	Standard Deviation
pH (pH units)	0.01	0.02
pCO <sub>2</sub> (mmHg)	1.35	0.96
pO <sub>2</sub> Arterial (>80 mmHg)	7.24	1.48
pO <sub>2</sub> Venous (<80 mmHg)	1.22	0.85
SO <sub>2</sub> (%)	0.69	0.85
Total Hgb (g/dL)	0.44	0.40
K <sup>+</sup> (mmol/L)	0.15	0.09
HCT Value	1.35	1.19

Flow Sensor Q		
1/4" x 1/16" (% of reading)	3.77	7.54
1/4" x 3/32" (% of reading)	4.26	6.76
3/8" x 3/32" (% of reading)	4.82	2.13

**Table 1:** Mean measurement error and standard deviation differences found.



**Figure 2:** Components of the CDI OneView Monitoring System.

## CDI OneView Monitoring System Technologies

### System Overview

The CDI OneView Monitoring System consists of a Central Processing Core, Touchscreen Display, a user-selected combination of Flow Sensors, Data Modules, Blood Parameter Modules (BPMs) and Hematocrit/Oxygen Saturation Probes (H/S Probes), disposable sterile Shunt Sensors and H/S Cuvettes, and a Calibrator.

The disposable Shunt Sensors and H/S Cuvettes are installed in the corresponding BPMs and H/S Probes at points in the circuit which will allow adequate exposure to blood (see Figure 1).

Clinicians select the combination of BPMs and H/S Probes depending on the parameters to be monitored (see Table 2). BPMs, which measure arterial and/or venous pH, pCO<sub>2</sub>, pO<sub>2</sub>, K<sup>+</sup>, and temperature use optical fluorescence technology in conjunction with the disposable Shunt Sensor. The CDI OneView System calculates the following important parameters: arterial oxygen saturation (SaO<sub>2</sub>), BE, HCO<sub>3</sub><sup>-</sup>, VO<sub>2</sub>, DO<sub>2</sub>, O<sub>2</sub>ER, CI, and AUC. The H/S Probes, which measure HCT, Hgb, and arterial and/or venous SO<sub>2</sub>, use optical reflectance technology in conjunction with the disposable H/S Cuvettes.

Parameters Measured	Components	Technology Utilized
pH, pCO <sub>2</sub> , pO <sub>2</sub> , K <sup>+</sup>	BPM w/ Shunt Sensor	Optical fluorescence
HCT, Hgb, SO <sub>2</sub>	H/S Probe w/ H/S Cuvette	Optical reflectance
Q Flow Rate	Flow Module w/ Flow Sensor	Ultrasonic

**Table 2:** The CDI OneView System allows calculation of the following important parameters:

BE, HCO<sub>3</sub><sup>-</sup>, VO<sub>2</sub>, VO<sub>2</sub>i, DO<sub>2</sub>, DO<sub>2</sub>i, CI, AUC. The system calculates SaO<sub>2</sub> and SvO<sub>2</sub> if not measured.


### Optical Fluorescence with the CDI Shunt Sensor

The CDI OneView Monitoring System uses optical fluorescence technology with the Shunt Sensor to measure pH, pCO<sub>2</sub>, pO<sub>2</sub>, and K<sup>+</sup> in blood. The Shunt Sensor contains four microsensors — one each for pH, pCO<sub>2</sub>, pO<sub>2</sub>, and K<sup>+</sup> — and a thermistor to measure temperature. Temperature is measured at the Shunt Sensor site and does not represent the patient's actual temperature. The microsensors are in direct contact with the blood, enabling rapid response time.

The Shunt Sensor can be placed in any arterial or venous shunt or purge line with continuous flow (see Figure 1). A minimum blood flow requirement of 35 mL/min is necessary for proper measurement.


During normal operation of the CDI OneView Monitoring System, light emitting diodes (LEDs) in the BPMs direct light pulses toward the microsensors, which contain fluorescent dyes (see Figure 3). As these pulses strike the microsensors, fluorescent light is emitted. The intensity of the fluorescent light will vary depending on the pH, pCO<sub>2</sub>, pO<sub>2</sub>, and K<sup>+</sup> in the blood. A photo detector in the Cable Head measures the intensity of the fluorescent light and converts it to numerical data which is presented on the Touchscreen Display.

The pH, pCO<sub>2</sub>, and pO<sub>2</sub> measurements are taken every second. The K<sup>+</sup> measurement is taken every six seconds.



**Figure 3:** Optical fluorescence: CDI Shunt Sensor and BPM.






*H/S Cuvettes.*

## Optical Reflectance with the CDI System H/S Cuvette

The CDI OneView Monitoring System uses optical reflectance technology with the H/S Probes to measure oxy- and deoxy- forms of hemoglobin.

The flow-through H/S Cuvette is placed directly in the CPB circuit. A window in the H/S Cuvette allows optical measurement without blood contact (see Figure 4).

LEDs in the H/S Probe direct light pulses of specific wavelength at the blood through the optical window in the H/S Cuvettes (see Figure 4). The intensity of the resulting reflections are analyzed (on the basis of the characteristic spectra of the oxy- and deoxy- forms of hemoglobin) to determine the oxygen saturation, hematocrit, and hemoglobin. These measurements are taken every 18 milliseconds and are displayed on the monitor's Touchscreen Display.



**Figure 4:** Optical reflectance: CDI H/S Cuvette and H/S Probes.

## CDI H/S Cuvette Blood Flow Rates

The CDI OneView Monitoring System should only be used where there is blood flow through the extracorporeal circuit. To perform accurately, reference the minimum and maximum blood flow rates (Q) as listed in the CDI H/S Cuvette Instructions for Use. Restoration of blood flow above the minimum through the H/S Cuvette will restore performance of the system.

## Ultrasonic Technology with the CDI Flow Sensor

The flow sensor uses ultrasonic “transit-time” technology to measure flow through a clamp-on sensor specifically designed for use with flexible plastic tubing. Ultrasonic transducers pass signals back and forth, alternately intersecting the flowing fluid in upstream and downstream directions. The transit-time it takes for the wave of ultrasound to travel from one transducer to the other is proportional to flow, and the difference between the upstream and downstream transit times is used to calculate the flow rate.



*Placing H/S Probe onto H/S Cuvette.*

## CDI OneView Monitoring System Calibration of the Microsensors



*Shunt Sensors in the BPM placed within the Calibrator.*

### pH, pCO<sub>2</sub>, pO<sub>2</sub> Sensors

Sensors for pH, pCO<sub>2</sub>, and pO<sub>2</sub> are calibrated using a two-point tonometered calibration system, similar to the system used to calibrate the electrodes in laboratory analyzers.


The calibration process uses the CDI Model 740 Calibrator and two canisters of calibration gases, Gas 1 and Gas 2. The calibration gases contain precise, defined levels of carbon dioxide (CO<sub>2</sub>) and oxygen (O<sub>2</sub>) gases (see Table 3). During calibration, the Shunt Sensors (attached to the BPM) are placed in the Calibrator, allowing the calibration gases to flow through the buffer solution contained in each Shunt Sensor. This exposes the microsensors to the gases with known CO<sub>2</sub> and O<sub>2</sub> values.

	Gas 1 Model CDI746	Gas 2 Model CDI747
pH	7.171 pH units	7.599 pH units
CO <sub>2</sub>	7.5% ± 0.1% (55.6 mmHg)	2.8% ± 0.1% (20.8 mmHg)
O <sub>2</sub>	24.0% ± 0.1% (177.9 mmHg)	4.0% ± 0.1% (29.7 mmHg)

**Table 3:** Calibrating gas values for the CDI OneView System.  
Balance of gas mixture is nitrogen (N<sub>2</sub>). Gases measured at 1 atm and 21 °C.

To perform the calibration, the system measures the fluorescent intensities emitted by a microsensor as it is exposed to Gas 1 and then Gas 2. It then plots the two

fluorescent measurements as a function of the predefined values of the calibration gases (see example for pO<sub>2</sub> in Figure 5). The system uses the two points to create a slope and a y-intercept for that parameter. During CPB, as the system measures the fluorescent intensity of the blood in the extracorporeal circuit, it uses the slope and intercept to extrapolate corresponding blood parameter values.



**Figure 5:** Example of a 2-point calibration for pO<sub>2</sub> sensor.

### K<sup>+</sup> Sensor

Calibration of the K<sup>+</sup> microsensor also relies on a two-point slope and intercept calibration process. The slope is defined using the factory-measured value encoded in the Shunt Sensor. This is the K<sup>+</sup> code on the Shunt Sensor pouch which is entered during the initial calibration sequence (as described in the CDI OneView Monitoring System Operator's Manual). The intercept point is obtained after the initiation of CPB using the K<sup>+</sup> level in a patient blood sample. The sample is drawn, the CDI OneView Monitoring System K<sup>+</sup> reading is stored in the system, the sample is processed using the laboratory analyzer, and the analyzer's value is entered into the CDI OneView Monitoring System to recalibrate the stored reading.

### H/S Probe

Each H/S Probe is precalibrated at the factory for SO<sub>2</sub>, HCT, and Hgb values; and further calibration is not required before going on CPB. A color chip test is performed at system startup. To meet system accuracy limits, perform an in vivo re-calibration as outlined in the CDI OneView Monitoring System Operator's Manual.



CDI OneView Monitoring System.

## Glossary of Acronyms and Abbreviations

<b>AKI</b>	Acute Kidney Injury	<b>HLM</b>	Heart-lung Machine
<b>AUC</b>	Area Under the DO <sub>2</sub> i Curve	<b>ICU</b>	Intensive Care Unit
<b>BE</b>	Base Excess	<b>K<sup>+</sup></b>	Potassium Ion
<b>BPM</b>	Blood Parameter Module	<b>O<sub>2</sub></b>	Oxygen
<b>BSA</b>	Body Surface Area	<b>O<sub>2</sub>ER</b>	Oxygen Extraction Ratio
<b>CI</b>	Cardiac Index	<b>paO<sub>2</sub></b>	Arterial Partial Pressure of Oxygen
<b>CPB</b>	Cardiopulmonary Bypass	<b>pCO<sub>2</sub></b>	Partial Pressure of Carbon Dioxide
<b>DMS</b>	Data Management System	<b>pH</b>	Potential of Hydrogen
<b>DO<sub>2</sub></b>	Oxygen Delivery	<b>pO<sub>2</sub></b>	Partial Pressure of Oxygen
<b>DO<sub>2</sub>i</b>	Indexed Oxygen Delivery	<b>Q</b>	Blood Flow Rate
<b>EMR</b>	Electronic Medical Records	<b>rSO<sub>2</sub></b>	Cerebral Regional Oxygen Saturation
<b>GDP</b>	Goal-Directed Perfusion	<b>SaO<sub>2</sub></b>	Arterial Oxygen Saturation
<b>H/S</b>	Hematocrit/Oxygen Saturation	<b>SO<sub>2</sub></b>	Oxygen Saturation
<b>HCO<sub>3</sub><sup>-</sup></b>	Bicarbonate	<b>SvO<sub>2</sub></b>	Venous Oxygen Saturation
<b>HCT</b>	Hematocrit	<b>VO<sub>2</sub></b>	Oxygen Consumption
<b>Hgb</b>	Hemoglobin	<b>VO<sub>2</sub>i</b>	Indexed Oxygen Consumption
<b>HL</b>	Hyperlactatemia		

## Product Specifications

Parameters	Display Range	Resolution
pH	6.50 to 8.50	0.01
pCO <sub>2</sub>	10 to 200 mmHg 1.3 to 26.7 kPa	1 0.1
pO <sub>2</sub>	10 to 700 mmHg 1.3 to 93.3 kPa	1 0.1
K <sup>+</sup>	1.0 to 9.9 mmol/L	0.1
Shunt Temp	1.0 to 45.0 °C	0.1
SO <sub>2</sub>	35 to 100%	1%
HCT	12 to 45%	1%
Hgb	4.0 to 15.0 g/dl	0.1
Q	0 to 10.00 L/min	0.001 < 1.00 L/min 0.01 ≥ 1.00 L/min
CI	0.1 to 10.0 L/min/m <sup>2</sup>	0.1
BE	-25 to 25 mEq/L	1
HCO <sub>3</sub> <sup>-</sup>	0 to 50 mEq/L	1
VO <sub>2</sub>	1 to 400 mL/min	1
VO <sub>2</sub> i	1 to 1000 mL/min/m <sup>2</sup>	1
DO <sub>2</sub>	1 to 3000 mL/min	1
DO <sub>2</sub> i	1 to 1000 mL/min/m <sup>2</sup>	1
rSO <sub>2</sub>	Match source device	Match source device
Q (from HLM)	Match source device	Match source device
O <sub>2</sub> ER	0.00 to 100%	1.00%
SaO <sub>2</sub> calc	35 to 100%	1%
SvO <sub>2</sub> calc	35 to 100%	1%

Component	Part #	Dimensions	Weight
Core	CDI750	22.2 cm x 27.2 cm x 8.3 cm	3 kg
Display + Backplate	CDI751	30 cm x 20.5 cm x 9.5 cm	2.4 kg
Calibrator	CDI740	23 cm x 21 cm x 18 cm	3.3 kg
Core Mounting Arm	CDI780	21 cm x 15.5 cm x 12 cm	1.6 kg
Display Mounting Arm	CDI781	47 cm x 28 cm x 9 cm	1.7 kg
BPM Mounting Arm	CDI782	32.4 cm x 18.6 cm x 8.2 cm	1 kg
Calibrator Mounting Arm	CDI783	19.2 cm x 9 cm x 5.8 cm	1 kg

### System Power Requirements and Specifications

100-240 VAC, 50/60 Hz

14.4-volt backup battery

### Model CDI510H Shunt Sensor

Sterile, heparin-treated

Priming volume 1.2 mL

### System Display Update\*

Every 1 second

### System Measurement Cycle Time\*

pH, pCO<sub>2</sub>, pO<sub>2</sub>, SO<sub>2</sub>, HCT, Hgb, Q = one measurement per second

K<sup>+</sup> = one measurement per six seconds

### CDI H/S Cuvette

Sterile

Priming volume for # 6914 (1/4" x 1/4") = 4 mL

Priming volume for # 6913 (3/8" x 3/8") = 9 mL

Priming volume for # 6912 (1/2" x 1/2") = 16 mL



#### Terumo Cardiovascular

6200 Jackson Road, Ann Arbor, MI 48103-9300 USA  
Tel: +1.734.663.4145  
Email: cardiovascular.emea@terumo-europe.com

For a complete list of offices, please consult our website.

[Terumo-europe.com](https://www.terumo-europe.com)

#### Terumo Corporation Tokyo Office

Toranomon Hills Station Tower 28F, 2-6-1  
Toranomon, Minato-ku Tokyo, 105-5528, Japan  
TEL +81-3-6742-8500

#### Terumo Europe N.V.

Authorized EC Representative  
Interleuvenlaan 40, B-3001 Leuven, Belgium  
Tel: +32.16.38.12.11

#### Terumo Asia Holdings Pte. Ltd.

300 Beach Road, #33-03, The Concourse Singapore 199555  
Tel: +65.6.295.1792

#### Terumo Latin America Corporation

121 Alhambra Plaza Suite 1120, Coral Gables, FL 33134  
Tel: +1.305.477.4822

Technology inspired for life.™

©2024 Terumo Europe N.V. CV434GB 06.24.SK-I(06.24)E. Company names and brand names used herein/hereon are trademarks or registered trademarks of TERUMO CORPORATION, its affiliates, or unrelated third parties. Refer to this device's Operator's Manual for the risks associated with its use.

\* Internal testing data on file.

CE-Mark Pending